

DEC 01 2009

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130404</u> ✓
A	Name <u>John W Carroll</u> ID# <u>6775</u> Date <u>11/28/09</u> ✓	
	Agency <u>Fort Greely Police BIG 4</u> Phone # <u>907-873-9134</u>	
	Instrument Location <u>Bld 663 Fort Greely, Alaska</u> ✓ <u>Booking Room</u>	
B	Alco S/N <u>X172958</u> Target Value <u>.087</u> High Pressure <u>1000</u>	
	Alco Test Value Average <u>.085</u> ✓ <u>.084</u> ✓ 1 st Alco 2 nd Alco	
	Signature <u>John W Carroll</u> ✓	
(OVER)		

BMB
12/31/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of January, 2010.
Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

NOVEMBER 28, 2009

OPERATOR'S NAME:

CARROLL, JOHN

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 000000

DEPT/AGENCY: FT GREELY PD B164

CASE/REPORT: 000000

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: X172958

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.41 in
ALCO TARGET .085 12:05
BLANK TEST .000 12:06
INTERNAL STANDARD VERIFIED 12:06
ALCO TV 29.41 in .085 12:06
BLANK TEST .000 12:07
SUBJECT SAMPLE .000 12:08
BLANK TEST .000 12:08
ALCO TV 29.41 in .084 12:09
BLANK TEST .000 12:09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

NOVEMBER 28, 2009

TIME 12:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.41 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~